

NDS Products, Inc.

111 Anderson Street
Pasadena, TX 77506

Phone: 713/475-2986 Fax: 713/477-6741

Web Site: ndsproducts.com

e-mail: sales@ndsproducts.com info@ndsproducts.com

ND-2000

ND-2000A

ND-2000C

ND-2000P

Operation Manual

Revision B November 03, 2005



CONTENTS

ND-2000 - ND-2000C – ND-2000A – ND-2000P

PARAGRAPH		PAGE
1.0	USER CAUTION AND WARNINGS	2
2.0	WARRANTY	2
3.0	MECHANICAL SPECIFICATIONS	3
4.0	ELECTRICAL SPECIFICATIONS	3
4.5	AURAL MONITORING	4
5.0	SURVEY METER LIMITATIONS (SATURATION)	5
6.0	INSTRUMENT OPERATION	5
6.1	CHECKOUT	5
6.2	RANGE SELECTION/ TAKING READINGS	5
6.3	CHARGING DOSIMETER	5
7.0	MAINTENANCE	6
7.5	CAUTION NOTE	6
8.0	CALIBRATION PROCEDURE	6
9.0	CURRENT ADJUSTMENT PROCEDURE	7
10.0	BATTERY CHECK ADJUSTMENT PROCEDURE	8
11.0	SERVICING AND TROUBLE-SHOOTING	9
12.0	CONVERSION TABLE	10
13.0	PARTS LIST	11
14.0	CIRCUIT DIAGRAM	12

NOTE: DUE TO NDS PRODUCTS' CONTINUING PROGRAM OF RESEARCH & DEVELOPMENT, ALL SPECIFICATIONS ARE SUBJECT TO CHANGE WITHOUT NOTICE AND MAY BE VARIED AT OUR DISCRETION.

1.0 USER CAUTION AND WARNINGS:

THIS INSTRUMENT IS INTENDED SOLELY FOR THE DETECTION AND MEASUREMENT OF IONIZING RADIATION. IT SHOULD BE USED ONLY BY PERSONS WHO HAVE BEEN TRAINED IN THE PROPER INTERPRETATIONS OF ITS READINGS AND THE APPROPRIATE SAFETY PROCEDURES TO BE FOLLOWED IN THE PRESENCE OF RADIATION.

ALL INSTRUCTIONS AND WARNINGS CONTAINED IN THIS MANUAL OR ON THE INSTRUMENT ITSELF, MUST BE READ BEFORE USE AND MUST BE STRICTLY FOLLOWED. FAILURE IN FOLLOWING THESE INSTRUCTIONS AND WARNINGS MAY RESULT IN INACCURATE READINGS AND/OR USER HAZARD. INDICATED BATTERY AND OTHER OPERATIONAL TESTS MUST BE PERFORMED PRIOR TO EACH USE TO ASSURE THAT THE INSTRUMENT IS FUNCTIONING PROPERLY.

FAILURE TO CONDUCT PERIODIC PERFORMANCE TESTS IN ACCORDANCE WITH ANSI N323-1978, 4.6, AND 5.4, AND TO KEEP RECORDS THEREOF IN ACCORDANCE WITH PARAGRAPH 4.5, COULD RESULT IN ERRONEOUS READINGS OF POTENTIAL DANGER. ANSI N323-1978 BECOMES, BY THIS REFERENCE, A PART OF THIS OPERATING MANUAL.

THE EQUIPMENT HEREIN IS DESIGNED AND MANUFACTURED IN COMPLIANCE WITH ALL APPLICABLE SAFETY STANDARDS; NEVERTHELESS, CERTAIN HAZARDS ARE INHERENT IN THE USE OF ALL ELECTRONIC EQUIPMENT. ADEQUATE WARNINGS ARE INCLUDED IN THIS MANUAL AND ON THE PRODUCT ITSELF TO COVER HAZARDS THAT MAY BE ENCOUNTERED IN NORMAL USE AND SERVICING OF THIS EQUIPMENT. NO OTHER PROCEDURES ARE WARRANTED BY NDS PRODUCTS. IT SHALL BE THE OWNER'S OR USER'S RESPONSIBILITY TO ASSURE THAT THE PROCEDURES HEREIN ARE METICULOUSLY FOLLOWED, AND ESPECIALLY THAT THE WARNING AND CAUTIONARY NOTES ARE HEEDDED. FAILURE ON THE PART OF THE USER IN ANY WAY TO FOLLOW PRESCRIBED PROCEDURES SHALL OBSOLVE NDS PRODUCTS AND ITS AGENTS FROM ANY RESULTING LIABILITY.

READ YOUR OPERATION MANUAL

2.0 WARRANTY:

NDS PRODUCTS WARRANTS THAT THE PRODUCTS COVERED HEREBY SHALL BE FREE FROM DEFECTS IN WORKMANSHIP AND MATERIALS FOR A PERIOD OF ONE YEAR FROM DATE OF NDS PRODUCTS' SHIPMENT (GM TUBE IS WARRANTED BY THE OEM FOR 90 DAYS). THE FOREGOING WARRANTY DOES NOT APPLY TO ANY PRODUCTS WHICH HAVE BEEN SUBJECT TO MISUSE (INCLUDING STATIC DISCHARGE), NEGLIGENCE, ACCIDENT, OR MODIFICATION. NDS PRODUCTS' SOLE OBLIGATION TO BUYER HEREUNDER FOR PRODUCTS FAILING TO MEET THE AFORESAID WARRANTY SHALL BE, AT NDS PRODUCTS' DISCRETION, TO REPAIR AND RECALIBRATE THE NONCONFORMING PRODUCT OR TO REPLACE THE NONCONFORMING PRODUCT WHERE WITHIN THE WARRANTY PERIOD: 1) THE BUYER HAS RETURNED THE NONCONFORMING PRODUCT TO NDS PRODUCTS, FREIGHT PREPAID, & 2) NDS PRODUCTS HAS DETERMINED THE PRODUCT IS NONCONFORMING AND THAT SUCH NONCONFORMITY IS NOT A RESULT OF IMPROPER USE, REPAIR, MODIFICATION OR OTHER MISUSE BY BUYER.

THE FOREGOING WARRANTY AND REMEDIES ARE EXCLUSIVE AND ARE MADE EXPRESSLY IN LIEU OF ALL OTHER WARRANTIES EXPRESSED OR IMPLIED, EITHER IN FACT OR BY OPERATION OF LAW, STATUTORY OR OTHERWISE, INCLUDING WARRANTIES OF MERCHANTABILITY AND FITNESS FOR USE. NDS PRODUCTS NEITHER ASSUMES NOR AUTHORIZES ANY OTHER PERSON TO ASSUME FOR IT ANY OTHER LIABILITY IN CONNECTION WITH THE SALE OR USE OF ITS PRODUCTS, AND NDS PRODUCTS MAKES NO WARRANTY WHATSOEVER FOR PRODUCTS NOT MANUFACTURED BY NDS PRODUCTS. NDS PRODUCTS SHALL NOT BE LIABLE FOR DAMAGES DUE TO DELAYS IN DELIVERIES OR USE AND SHALL IN NO EVENT BE LIABLE FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES OF ANY KIND; WHETHER ARISING FROM CONTRACT, TORT, OR NEGLIGENCE, INCLUDING BUT NOT LIMITED TO, LOSS OF PROFITS, LOSS OF GOODWILL, OVERHEAD OR OTHER LIKE DAMAGES.

3.0 MECHANICAL SPECIFICATIONS:

THE MODEL ND-2000 SERIES IS A SENSITIVE, PORTABLE PULSE COUNT RATEMETER AND POWER SUPPLY WHICH WILL DETECT GAMMA AND X-RAY RADIATION. THE INSTRUMENT UTILIZES THE LATEST DESIGN IN SOLID STATE CIRCUITRY AND CONTAINS NO VACUUM TUBES RESULTING IN A RUGGED AND RELIABLE DEVICE. THIS DESIGN ALLOWS THE INSTRUMENT TO OPERATE ON ONLY 2 "D" CELL BATTERIES. IT PROVIDES A BATTERY LIFE OF OVER 150 HOURS WITH CONTINUOUS OPERATION AND LONGER WITH INTERMITTENT USE (THE ABOVE IS TRUE WITH NON-ALKALINE BATTERIES, LONGER WITH ALKALINE BATTERIES).

ND-2000 SERIES FEATURES: A RUGGED, TWO PIECE, NYLON REINFORCED POLYESTER RESIN CASE WITH A NEOPRENE O-RING CASE GASKET; A ROTARY SHAFT SEAL; NEOPRENE GASKETS FOR THE METER INDICATOR; RUGGED METAL HOUSING METER INDICATOR; AND A DOSIMETER CHARGER CONTACT WITH ZEROING POTENTIOMETER (ND-2000C ONLY).

4.0 ELECTRICAL SPECIFICATIONS:

4.1 POWER SUPPLY:

THE HIGH VOLTAGE POWER SUPPLY IS A BLOCKING OSCILLATOR DRIVEN "FLY-BACK" TYPE CIRCUIT. THE BLOCKING OSCILLATOR PORTION OF THE CIRCUIT CONSISTS OF TRANSISTOR Q3, WINDINGS 3-4 AND 5-6 OF TRANSFORMER T1, POTENTIOMETER R17 AND BATTERIES. C8 SERVES ONLY TO SUPPRESS HIGH FREQUENCY PARASITIC OSCILLATION CAUSED BY THE TRANSISTOR PARAMETERS. THE HIGH VOLTAGE POWER SUPPLY PORTION OF THE CIRCUIT CONSISTS OF WINDINGS 1-2 OF T1, RECTIFIER CR5, THE ASSOCIATED RESISTORS AND CAPACITORS, AND DIODES Z1 THROUGH 4. THE LOW VOLTAGE SECTION IS COMPRISED OF WINDINGS 5-6 OF T1, DIODE CR4 AND C7. THE OPERATION OF THE POWER SUPPLY IS AS FOLLOWS: WHEN THE INSTRUMENT IS ON, Q3 CONDUCTS WHILE AN INCREASING CURRENT FLOWS THROUGH WINDING 3-4 AND THE COLLECTOR OF Q3. THIS CURRENT INDUCES A VOLTAGE IN WINDING 5-6 OF SUCH POLARITY AS TO SUSTAIN AND INCREASE THE CONDUCTION OF Q3. THE COLLECTOR CURRENT CONTINUES TO INCREASE UNTIL Q3 & WINDING 3-4 BECOMES CONSTANT. WHEN THE CURRENT IN WINDING 3-4 BECOMES CONSTANT, THE INDUCED VOLTAGE IN WINDING 5-6 FALLS TO ZERO WHICH CAUSES THE BASE CURRENT TO DROP, WHICH IN TURN, CAUSES THE CURRENT FLOWING THRU THE COLLECTOR AND WINDING 3-4 TO DROP. THIS DECREASING CURRENT INDUCES A VOLTAGE IN WINDING 5-6 OF SUCH POLARITY AS TO TURN OFF THE TRANSISTOR. INsofar AS THIS IS A

REGENERATIVE ACTION, Q3 TURNS OFF EXTREMELY FAST, CAUSING THE FLUX IN T1 TO COLLAPSE SUDDENLY; THIS LATER IS THE "FLY-BACK" ACTION, WHICH INDUCES HIGH VOLTAGE ON ALL WINDINGS. THE MAGNITUDE OF THE VOLTAGE IS PROPORTIONAL TO THE NUMBER OF TURNS ON THE WINDING. THE VOLTAGE INDUCED ON WINDING 1-2 IS EXTREMELY HIGH DUE TO THE LARGE NUMBER OF TURNS. THIS IS RECTIFIED BY CR5 AND FILTERED BY C9 AND R13, AFTER WHICH IT IS REGULATED TO 600V BY Z1 THROUGH Z4. THE REGULATING ACTION OF THE ZENERS IS REFLECTED BACK THROUGH T1, AND THUS THE VOLTAGE INDUCED AT THE OTHER WINDING IS ALSO REGULATED. THIS IS TAKEN ADVANTAGE OF AT WINDING 5-6 WHERE THE INDUCED VOLTAGE IS RECTIFIED BY CR4 AND USED TO POWER THE REST OF THE INSTRUMENT. THE INDUCED VOLTAGE AT THE BASE THEN RETURNS TO ZERO, ALLOWING Q3 TO CONDUCT AGAIN AND THUS REPEATING THE CYCLE; THE RATE OF REPETITION BEING CONTROLLED BY R17.

4.2 MONOSTABLE MULTIVIBRATOR:

THE MONOSTABLE MULTIVIBRATOR CIRCUIT CONSISTS OF Q1, Q2, CALIBRATION POTS R3-5, R20, R26, S1, AND ASSOCIATED COMPONENTS. ITS FUNCTION IS TO PROVIDE A UNIFORM CURRENT PULSE OUTPUT FOR EACH PULSE INPUT, REGARDLESS OF THE SHAPE OR MAGNITUDE OF THE INPUT PULSE.

4.3 METER AND TIME CONSTANT CIRCUIT:

THE METERING CIRCUIT CONSISTS OF M1 (METER), R7, AND C2 (THE INTEGRATING CAPACITOR). WHEN A PULSE CAUSES Q1 TO CONDUCT, THE COLLECTOR CURRENT PASSES THRU C2 LEAVING IT CHARGED. IT IS THEN DISCHARGED THROUGH M1 CAUSING IT TO DEFLECT. THE AMOUNT OF DEFLECTION IS PROPORTIONAL TO THE AMOUNT OF CHARGE, WHICH IN TURN IS PROPORTIONAL TO THE AVERAGE CURRENT. THUS M1 READS THE AVERAGE CURRENT THROUGH Q1, WHICH IS PROPORTIONAL TO RATE AND WIDTH. THE RESPONSE TIME OF THE SYSTEM IS A FUNCTION OF THE SIZE OF CAPACITOR C2. THE LARGER THE CAPACITOR THE LONGER IT TAKES TO CHARGE AND DISCHARGE, THUS THE RESPONSE TIME CAN BE ALTERED BY CHANGING THE AMOUNT OF CAPACITY IN THE CIRCUIT.

4.4 DETECTOR:

THE DETECTOR USED IN THIS INSTRUMENT IS AN ENERGY COMPENSATED, Ne/HALOGEN GAS FILLED GEIGER MUELLER TUBE. THE EFFECTIVE LENGTH IS 0.625 INCH WITH AN EFFECTIVE DIAMETER OF 0.194 INCH. THE WALL THICKNESS IS 30 mg/cm SQUARED.

4.5 OPTIONAL AURAL MONITOR:

THIS ASSEMBLY CONSISTS OF AN AUDIO TRANSDUCER AND ITS ASSOCIATED COMPONENTS. THE OPERATOR WILL NOTE THAT EACH PULSE ARRIVING AT THE INSTRUMENT IS INDICATED BY A DISTINCTLY AUDIBLE "CLICK" FROM THE TRANSDUCER. THESE CLICKS OR PULSES, AS THEY ARE PRODUCED FROM A RADIATION DETECTOR TUBE, ARE RANDOMLY SPACED IN TIME SO THAT ONE MAY WAIT FOR SEVERAL SECONDS BEFORE ANY CLICKS ARE HEARD AND THEN THERE MAY BE TWO OR THREE IN RAPID SUCCESSION. ACCURATE MEASUREMENTS OF BACKGROUND AND OTHER

LOW LEVEL RADIATION CAN BE MADE BY COUNTING AND TIMING THE SPEAKER CLICKS WITH A WATCH HAVING A SECOND HAND.

5.0 SURVEY METER LIMITATION (SATURATION) *CAUTION*

IT IS A WELL KNOWN PHENOMENON THAT ALL GEIGER-MUELLER SURVEY METERS SATURATE OR OVERLOAD AT SOME POINT (THIS MAY BE CALLED FLOODING OR JAMMING). THIS MODEL WILL SATURATE WHEN IN A RADIATION FIELD IN EXCESS OF 1000 R/hr WHEN IT HAS BEEN PROPERLY MAINTAINED. IF THE RADIATION FIELD IS IN EXCESS OF 1000 R/hr ONE OF TWO EVENTS WILL OCCUR: 1) THE NEEDLE WILL SWING TO THE EXTREME RIGHT AND WHEN IT REACHES SATURATION, THE NEEDLE WILL FALL BACK TO ZERO, OR 2) IF THE SURVEY METER IS TURNED ON IN A FIELD IN EXCESS OF 1000 R/hr, THE NEEDLE WILL NOT REACT AND STAY AT ZERO.

ALTHOUGH A 1000 R/hr FIELD IS EXTREMELY HIGH AND MAY NEVER BE ENCOUNTERED, PROPER RADIATION SAFETY PROCEDURES AND REGULATIONS SHOULD BE OBSERVED WHEN MAKING ANY TYPE OF RADIATION SURVEY.

6.0 INSTRUMENT OPERATION:

6.1 CHECK THE FOLLOWING BEFORE USING:

- 6.1.1 TURN SWITCH TO BATTERY CHECK. THE METER SHOULD READ IN THE INDICATED BATT ZONE (6.8-8.2). REPLACE BATTERIES IF NEEDED. DO NOT USE INSTRUMENT IF METER DIAL IS NOT REGISTERING IN BATT ZONE.
- 6.1.2 INSPECT FOR PHYSICAL DAMAGE.
- 6.1.3 CHECK CALIBRATION DUE DATE. DO NOT USE IF PAST DUE.
- 6.1.4 USE A CHECK SOURCE OR EQUIVALENT TO ENSURE THAT THE SURVEY METER IS READING RADIATION.
- 6.1.5 WHEN SECURING THE SURVEY METER LID TO CASE, CLEAN THE CASE GASKET TO ENSURE A PROPER SEAL. TIGHTEN THE 4 SCREWS UNTIL THE TOP AND BOTTOM MEET SECURELY. REPLACE THE GASKET WHEN NECESSARY.

6.2 SELECTING RANGE AND TAKING A READING:

TURN THE INSTRUMENT ON TO THE BATT CHECK POSITION AND SEE THAT THE METER READS IN THE INDICATED ZONE. WITH THE INSTRUMENT ON THE X100 (0-1000 mR/hr) RANGE, PLACE THE INSTRUMENT (PROBE) IN THE LOCATION TO BE MEASURED. IF THE READING IS LESS THAN 10% OF FULL SCALE, SWITCH TO THE X10 (0-100 mR/hr) RANGE. IF THE READING CONTINUES TO BE LESS THAN 10% OF FULL SCALE, CONTINUE SWITCHING TO MORE SENSITIVE RANGES: X1 (0-10 mR/hr) BEING THE MOST SENSITIVE. THIS RANGE IS USEFUL IN ESTABLISHING SAFE PERIMETERS. THE METER READING SHOULD ALWAYS BE MULTIPLIED BY THE RANGE SWITCH POSITION.

6.3 TO CHARGE (ZERO) DOISMETER: (ND-2000C ONLY)

- 6.3.1 TURN THE FUNCTION SWITCH TO THE BATT CHECK POSITION.

- 6.3.2 IF THE DOSIMETER HAS A PROTECTIVE CONTACT CAP, REMOVE IT. PRESS THE DOSIMETER CONTACT LIGHTLY INTO THE CHARGER CONTACT TO ILLUMINATE THE DOSIMETER SCALE.
- 6.3.3 PRESS THE DOSIMETER FIRMLY UNTIL THE HAIRLINE RESPONDS TO ADJUSTMENT OF THE ZEROING KNOB, AND ADJUST HAIRLINE TO ZERO. THE HAIRLINE MAY HAVE TO BE PLACED TO THE LEFT OF ZERO TO COMPENSATE FOR A KICK TO THE RIGHT WHEN THE DOSIMETER IS REMOVED. ALWAYS REMOVE DOSIMETER SLOWLY BY RELAXING THE PRESSURE FROM THE CHARGER.
- 6.3.4 TO READ THE DOSIMETER, POINT THE CONTACT END TOWARD AN EXTERNAL LIGHT SOURCE AND LOOK THROUGH THE EYEPEICE. IF NO LIGHT SOURCE IS AVAILABLE, PLACE THE DOSIMETER ON CHARGING CONTACT AND PRESS LIGHTLY TO ILLUMINATE THE SCALE.

7.0 MAINTENANCE:

- 7.1 BEFORE OPENING INSTRUMENT CASE FOR ANY REASON, BE SURE INSTRUMENT IS TURNED OFF.
- 7.2 KEEP CIRCUIT BOARD, SWITCH, BATTERY TERMINALS, AND GASKETS FREE OF OIL, MOISTURE, DIRT, BATTERY ACID, ETC.
- 7.3 CLEAN THE BATTERY TERMINAL CONTACTS WHEN INSTALLING NEW BATTERIES. USE FINE SAND PAPER OR EQUIVALENT TO ENSURE PROPER CONTACT.
- 7.4 WHEN CLEANING THE CIRCUIT BOARD AND SWITCH, USE A NON-RESIDUE CONTACT CLEANER. LET THE COMPONENTS DRY BEFORE CLOSING CASE OR USING INSTRUMENT.
- 7.5 *CAUTION* WHEN INSTALLING BATTERIES, NEVER TOUCH THE COMPONENTS OR CIRCUITRY. NEVER ADJUST POTENTIOMETERS R3, 4, 5, 17, 22, AND R26. THESE POTENTIOMETERS ADJUST CALIBRATION, CURRENT AND BATT CHECK READINGS. ADJUSTMENT MUST BE MADE BY QUALIFIED PERSONNEL WHO ARE FAMILIAR WITH CORECT PROCEDURES. INCORRECT ADJUSTMENTS WILL SEVERLY LIMIT THE SURVEY METER'S PERFORMANCE.
- 7.6 REMOVE BATTERIES WHEN INSTRUMENT IS TO BE STORED.


8.0 CALIBRATION PROCEDURES:

- 8.1 INSPECT CIRCUIT BOARD AND CASE FOR DAMAGE. CLEAN AND REPAIR IF NECESSARY.
- 8.2 INSPECT MECHANICAL HARDWARE AND TIGHTEN AS NECESSARY.
- 8.3 ZERO METER MOVEMENT AND CHECK GEOTROPISM.
- 8.4 CLEAN BATTERY CONTACTS.
- 8.5 INSTALL FRESH BATTERIES. TURN SURVEY METER TO BATT CHECK AND VERIFY THAT NEEDLE FALLS IN THE INDICATED BATT ZONE (6.8 TO 8.2).

- 8.6 CHECK THE HIGH VOLTAGE AT THE GM TUBE ANODE (540-600 VOLTS).
- 8.7 CHECK THE LOW VOLTAGE AT THE METER MOVEMENT (14-22 VOLTS).
- 8.8 CHECK THE CURRENT DRAW (17-22 mA).
- 8.9 THERE ARE 3 CALIBRATION POTENTIOMETERS LOCATED ON THE RIGHT SIDE OF THE CIRCUIT BOARD, FROM LEFT TO RIGHT: X100, X10, AND X1.
- 8.10 TO CALIBRATE, PLACE THE INSTRUMENT (OR PROBE) IN A KNOWN RADIATION FIELD FOR THE THREE RANGES AND WHEN NECESSARY, ADJUST THE CORRESPONDING POTENTIOMETER.
- 8.11 AFTER ADJUSTMENT, CHECK THE CALIBRATION NEAR THE ENDS OF EACH RANGE, AT APPROXIMATELY 20% AND 80% OF FULL SCALE. INSTRUMENT READINGS SHALL BE WITHIN +/- 20% OF KNOWN RADIATION VALUES AT THESE TWO POINTS OF EACH RANGE.

NOTE: THE CALIBRATION SOURCE MUST CONSTITUTE THE SOLE SOURCE OF RADIATION WHEN CALIBRATION IS PERFORMED. CALIBRATION MUST NOT BE UNDERTAKEN WHEN BACKGROUND IS ABOVE NORMAL OR WHEN THE INSTRUMENT IS IN A RADIOACTIVE FIELD OTHER THAN THAT PRODUCED BY THE KNOWN CALIBRATION SOURCE USED.

9.0 CURRENT ADJUSTMENT PROCEDURE:



- 9.1 ASSURE INSTRUMENT IS TURNED OFF.
- 9.2 CLEAN CIRCUIT BOARD.
- 9.3 CLEAN BATTERY TERMINAL CONTACTS WITH FINE SANDPAPER.
- 9.4 INSTALL FRESH BATTERIES.
- 9.5 CHECK HIGH AND LOW VOLTAGES:
- HIGH VOLTAGE: 540-600V AT GM TUBE.
 - LOW VOLTAGE: 14-22V AT METER MOVEMENT.
- 9.6 HOOK UP SURVEY METER AS SPECIFIED IN ABOVE DIAGRAM.

9.7 TURN SURVEY METER ON AND ADJUST R17 TO 17-22 ma.

9.8 AFTER SETTING CURRENT, ROTATE RANGE SWITCH AND CHECK THE CURRENT ON EACH RANGE. IF THE CURRENT IS CHANGING MORE THAN 10%, CLEAN OR REPLACE RANGE SWITCH. REMOVE BATTERIES PRIOR TO CLEANING.

10.0 BATTERY CHECK ADJUSTMENT PROCEDURE:

10.1 FOLLOW CURRENT ADJUSTMENT PROCEDURES STEP 1, 2, & 3.

10.2 TURN SURVEY METER ON AND LET THE SURVEY METER WARM-UP FOR ABOUT 2 MINUTES. PRESS BATTERY CHECK SWITCH.

10.3 ADJUST R22 UNTIL THE NEEDLE IS AT THE RIGHT HAND EDGE OF THE INDICATED BATT ZONE ON THE METER MOVEMENT (8.2).

WARNING: DO NOT USE R17 TO ADJUST THE BATTERY CHECK POSITION.

11.0 SERVICING AND TROUBLE-SHOOTING, SURVEY METER:

11.1 USE ONLY REPLACEMENT PARTS LISTED ON THE PARTS LIST AND CIRCUIT DIAGRAM.

11.2 WHEN REPLACING THE METER MOVEMENT, RTV RUBBER SILICONE MUST BE INJECTED INTO THE METER MOVEMENT MOUNTING HOLES. PLACE METER MOVEMENT INTO INSTRUMENT CASE, SECURE WITH HARDWARE AND REMOVE EXCESS RTV. THIS MUST BE PERFORMED TO ENSURE WATER TIGHTNESS.

11.3 THE GM TUBE IS THE MAJOR CAUSE IN SURVEY METER FAILURE. SOME OF THE PROBLEMS A DEFECTIVE TUBE CAN CREATE ARE AS FOLLOWS:

- CRACKED ANODE = DEAD TUBE, WILL NOT DETECT RADIATION OR WILL SATURATE EARLY.
- HIGH BACKGROUND = GAS CONTENT IS DEPLETING; TUBE WILL EVENTUALLY CEASE TO FUNCTION.
- GM TUBE SHORT = SURVEY METER'S INDICATOR WILL PEG OUT OR SATURATE EARLY.
- EARLY SATURATION = GM TUBE WILL DETECT RADIATION TO A CERTAIN POINT, i.e. 500 mR/hr, AND THEN BEGIN TO FALL DOWN SCALE. SOME TUBES MAY SHORT OUT AND THEN FALL DOWN SCALE.

NOTE: ROUTINE CALIBRATION MAY NOT ALWAYS DETECT THAT A GM TUBE IS SATURATING EARLY. SOME TUBES CAN SATURATE AT ANYTIME, THEREFORE IT IS HIGHLY RECOMMENDED THAT EACH TIME A SURVEY METER IS CALIBRATED; IT SHOULD BE TESTED IN A RADIATION FIELD OF 10 R/hr.

ALL NDS SURVEY METERS ARE TESTED FOR SATURATION AT OUR FACILITIES AFTER EACH CALIBRATION AND OR REPAIR.

11.4 DOSIMETER CHARGER: (ND-2000C ONLY)

11.4.1 IF THE LIGHT IS WEAK OR FAILS TO GO ON WHEN THE CONTACT IS DEPRESSED:

- REPLACE BATTERIES.
- CLEAN BATTERY AND LIGHT SWITCH CONTACTS.
- TIGHTEN OR REPLACE BULB.

11.4.2 HAIRLINE MOVES FLUIDLY BUT WILL NOT ZERO:

- REPLACE BATTERIES.

11.4.3 HAIRLINE MOVES ERRATICALLY:

- CLEAN CHARGING CONTACTS ON CHARGER AND DOSIMETER.
- REPAIR CHARGER.
- FAULTY DOSIMETER.

NOTE: ALWAYS KEEP THE CHARGER AND ESPECIALLY THE CHARGING CONTACT CLEAN BY USING A SOFT CLOTH. DO NOT USE A SOLVENT OR CLEANING FLUID ON PLASTIC CONTACT. ALWAYS KEEP THE PROTECTIVE CAP ON CHARGING CONTACT WHEN NOT IN USE.

Conversion Table mSv/hr AND uSv/hr

THIS MANUAL REFERENCES MILLI ROENTGENS PER HOUR (mR/hr), PLEASE NOTE THE FOLLOWING CONVERSIONS TO MILLI SIEVERTS PER HOUR (mSv/hr) AND MICRO SIEVERTS PER HOUR (uSv/hr) FOR THE

<u>mR/hr</u>	<u>mSv/hr</u>	<u>uSv/hr</u>
10 mR/hr	0.1 mSv/hr	100 uSv/hr
100 mR/hr	1 mSv/hr	1000 uSv/hr
1000 mR/hr	10 mSv/hr	10,000 uSv/hr

BILL OF MATERIAL

980-2000-000		ND-2000/C/A PC ASSY			REV D
ITEM	PART NUMBER	DESCRIPTION		QTY	REF DESIG
1	100-1000-222	CAP CER DISC	.0022 uF 1 KV 20%	1	C6
OPT	100-1000-472	CAP CER DISC	.0047 uF 1 KV 20%	1	C12
3	100-1000-103	CAP CER DISC	.01 uF 1 KV 20%	1	C9
4	100-0100-103	CAP CER DISC	.01 uF 100V 20%	1	C8
5	100-1000-223	CAP CER DISC	.022 uF 1 KV 20%	1	C17
6	105-0035-685	CAP ELE ALUM	6.8 uF 35V 20%	2	C1, 10
7	105-0025-226	CAP ELE ALUM	22.0 uF 35V 20%	1	C7
8	105-0006-477	CAP ELE ALUM	470.0 uF 25V 20%	1	C2
9	115-0100-222	CAP PLY FILM	.0022 uF 100V 10%	1	C3
10	115-0100-223	CAP PLY FILM	.022 uF 100V 10%	1	C5
12	150-0100-001	DIODE 1N4007	STANDARD RECOVERY	1	CR5
13	150-0200-001	DIODE 1N457A		4	CR1-4
OPT	150-0350-001	DIODE 1N5262B	ZENER	1	Z3
15	150-0400-001	DIODE 1N6031B	ZENER	3	Z1, 4, 6
16	250-1500-185	RES TF 1/2W 5%	1.8 M OHM 1KV 350PPM	1	R13
17	250-0100-222	RES CF 1/4W 5%	2.2 K OHM	2	R6-7
18	250-0100-335	RES CF 1/4W 5%	3.3 M OHM	1	R19
19	250-0100-103	RES CF 1/4W 5%	10.0 K OHM	2	R1, 2
20	250-0100-183	RES CF 1/4W 5%	18.0 K OHM	1	R10
21	250-0100-393	RES CF 1/4W 5%	39.0 K OHM	1	R11
22	250-0100-470	RES CF 1/4W 5%	47.0 OHM	1	R16
24	250-0100-154	RES CF 1/4W 5%	150.0 K OHM	1	R8
25	250-0100-334	RES CF 1/4W 5%	330.0 K OHM	2	R12, 21
OPT	250-0100-514	RES CF 1/4W 5%	510.0 K OHM	1	R29
28	265-0200-252	RES POT TRIMMER	2.5 K 1/2W	1	R17
30	265-0200-253	RES POT TRIMMER	25.0 K 1/2W	1	R5
31	265-0200-254	RES POT TRIMMER	250.0 K 1/2W	3	R3-4, 22
32	450-8050-001	TRANSFORMER	SWITCHING	1	T1
33	475-0150-001	TRANSISTOR	2N5086 PNP	2	Q1-2
34	475-0200-001	TRANSISTOR	SK3004 POWER	1	Q3
35	500-0714-001	TUBE GM	714 FOR 2000/2000A	1	V1

990-2000-000		ND-2000/C/A/P UNIT ASSY			REV B
ITEM	PART NUMBER	DESCRIPTION		QTY	REF DESIG
1	600-0095-001	BATTERY	1.5V D CELL EN95 ALK	2	BT 1-2
3	670-2000-001	CASE	NYLON/FIBERGLASS SET	1	
OPT	720-0100-001	CHARGER ASSY	CONTACT/ CAP	1	E1
OPT	690-2000-	DECAL ND-2000	XXX	1	
6	675-2000-002	HANDLE		1	
22	690-0100-001	KNOB	BLK/ ALUM INLAY	2	
OPT	575-0100-001	LAMP MINI	3V NO. 14	1	DS1
24	800-0010-001	METER 50uA	RUGGEDIZED METAL	1	M1
25	805-0100-001	METER GASKET		1	
OPT	395-2200-001	BOOT TOGGLE SWITCH		1	XS4
OPT	390-2200-001	SWITCH TOGGLE		1	S4
32	350-0300-004	SWITCH ROTARY	3 RANGE 5 POS	1	S1
33	355-0300-001	SWITCH ROTARY	SHAFT SEAL	1	
OPT	400-0200-001	TRANSDUCER ASSY		1	LS1
OPT	260-0100-506	RES POT CONTROL	5 MEG 10%/ 20% 2W	1	R23
OPT	SG101-07	PROBE SIDE WALL	FOR 2000P ONLY	1	V1

